# HITEMP140-TSK HITEMP140 DATA LOGGER WITH THERMAL SHIELD



#### Features

- Withstands Temperatures between
   -200 °C up to +250 °C
- Small Diameter: 2.0 in (51 mm)
- Submersible
- · Immediate or Delay Start
- Up to 1 Second Reading Rate

#### Benefits

- Validate a Wide Range of Temperature Processes Using One Data Logger
- HiTemp140 can be used With or Without Thermal Shield
- Durable Thermal Shield Protects
   Probe and Allows for a Fast

   Response Time

#### **Applications**

- · Peanut Roasting
- Food Processing
- · Meat Processing
- Autoclave Validation
- Conveyer Ovens
- Dishwasher Testing
- Incubator Validation

The HiTemp140-TSK is a kit that includes a HiTemp140 data logger with either a 5.25 inch probe or a 7 inch probe style, housed in a thermal shield. The combined features of the +0.1 % accuracy

of the HiTemp140 and the properties of the durable thermal shield allow the device to be used for a wide range of validation applications. This rugged system can be placed in and withstand temperature from -200  $^{\circ}$ C to +250  $^{\circ}$ C, making it ideal for use in autoclave validation, monitoring food processing and dishwasher testing.


Using the MadgeTech 4 software, the data logger is fast and easy to setup. Remove the thermal shield and place the HiTemp140 into the IFC400 or IFC406 docking station (sold separately). Using the software, an immediate or delay start can be chosen, as well as the reading rate. Select Start to program the settings and start the data logger. Place the thermal shield around the HiTemp140 and screw it back together. The device is ready to be deployed.

The HiTemp140-TSK can be completely submerged and is built for applications that require extreme temperature monitoring.

The HiTemp140-TSK flush style is designed to have the probe entirely exposed while the data logger is protected by the thermal sheild. This allows full use of the length of the probe for applications that require internal temperature monitoring. The Vented style offers more probe protection and is designed for shorter probe lengths in applications where the data logger might be subject to movement in a fully submerged application.



## MADGETECH DATA LOGGER SOFTWARE



#### Software Features:

- Multiple graph overlay
- Statistics
- Digital calibration
- Zoom in/zoom out
- Lethality equations (F0, PU)
- Mean Kinetic Temperature
- Full time zone support
- Data annotation
- Min./Max./Average lines
- Data table view
- Automatic report generation
- Summary view
- Multilingual

| Temperature Sensor:      | 100 Ω Platinum RTD  |  |  |
|--------------------------|---|--|--|
| Probe Measurement Range: | -200 °C to +260 °C (-328 °F to +500 °F)   |  |  |
| Temperature Resolution:  | 0.01 °C (0.02 °F)   |  |  |
| Calibrated Accuracy:     | • ±0.1 °C/±0.18 °F<br>(20 °C to +140 °C/68 °F to +284 °F)<br>• ±0.3 °C/±0.54 °F<br>(-20 °C to +19.99 °C/-4 °F to +67.98 °F)<br>• ±0.4 °C/±0.72 °F<br>(-40 °C to -20.01 °C/-40 °F to -4.02 °F) |  |  |
| Start Modes:             | Software programmable immediate start     Delay start up to 18 months in advance  |  |  |
| Stop Modes:              | Manual or Timed (specific date and time)  |  |  |
| Real Time Recording:     | May be used with PC to monitor and record data in real time   |  |  |
| Password Protection:     | An optional password may be programmed into the device to restrict access to configuration options. Data may be read out without the password.  |  |  |
| Memory:                  | 32,700 readings   |  |  |
| Wrap Around:             | Yes   |  |  |
| Reading Rate:            | 1 reading every second up to 1 reading every 24 hours   |  |  |

| Battery Type:                      | 3.6V high-temperature lithium battery included; user replaceable  |
|------------------------------------|---|
| Battery Life:                      | 1 year typical (1 minute reading rate at 25 °C/77 °F)   |
| Calibration:                       | Digital calibration through software  |
| Calibration Date:                  | Automatically recorded within device  |
| Data Format:                       | Date and time stamped °C, °F, K, °R   |
| Time Accuracy:                     | ±1 minute/month at 20 °C to 30 °C (68 °F to 86 °F)<br>(Stand alone mode)  |
| Computer Interface:                | IFC400 or IFC406 USB docking station required; 125,000 baud   |
| Software:                          | XP SP3/Vista/Windows 7/Windows 8 (MadgeTech 4 Only)   |
| Operating Environment:             | -40 °C to +140 °C (-40 °F to +284 °F), 0 %RH to 100 %RH   |
| Dimensions (Body):                 | 1.9 in x 0.97 in dia. (48 mm x 24.6 mm dia.)  |
| Model Number:                      | Dimensions (Probe)  |
| HITEMP40-5.25<br>HITEMP140-5.25-TD | 5.25 in x 0.188 in dia. (133 mm x 4.8 mm dia.)<br>5.25 in x 0.125 in dia. (0.188 in transitional dia.)<br>133 mm x 3.2 mm dia. (4.8 mm transitional dia.) |
| HITEMP140-7                        | 7.0 in x 0.188 in dia. (178 mm x 4.8 mm dia.)   |
| Weight:                            | 4.2 oz (120 g)  |
| Material:                          | 316 Stainless Steel   |
| Approvals:                         | CE  |

## HITEMP140-TSK SPECIFICATIONS\*\*

| Operating<br>Environment: | -200 °Cto +250 °C<br>(-328 °F to +482 °F) (Time limited)<br>0 %RH to 100 %RH   |
|---------------------------|--|
| Dimensions:               | <ul> <li>Flush Top: 2.75 in x 2.0 in dia.<br/>(69.85 mm x 51 mm dia.)</li> <li>Vented Top: 4.3 in x 2.0 in dia.<br/>(109.2 mm x 50.8 mm dia.)</li> </ul> |
| Material:                 | Enclosure: PTFE  |
| Weight:                   | Flush: 6.7 oz (190 g)     (not including data logger)     Vented: 9.5 oz (270 g)     (not including data logger)   |

## Disclaimer and Terms of Use

Listed specifications can be used to determine maximum allowable exposure times for the HiTemp140 with Thermal Shield at different temperatures beyond the normal operating range of the logger. Both the data logger and Thermal Shield must be at ambient temperature (approximately 25 °C) before being placed in the extreme temperature environment.

Immediately following exposure to high temperature, the data logger should be removed from the thermal shield (using appropriate precautions, as it could be VERY hot) OR the data logger and shield should be placed in a water bath (approximately 25 °C) for at least 15 minutes to allow it to cool. Failing to do this may allow heat trapped in the Thermal Shield to continue to heat the datalogger to potentially unsafe levels.

If your application involves a ramp up to a temperature above 140 °C and/or any complex temperature profile that isn't simply a constant temperature, please contact Madgelech to determine whether the HiTemp140 with Thermal Shield is suitable.

Please provide MadgeTech with a detailed description of your temperature profile, including temperatures, durations, ramp times, and process media (air, steam, oil, water, etc.) If MadgeTech is unable to definitively calculate the suitability of our product for your application, we can provide a test unit outfitted with a high temperature indicator sticker. This sticker has an indicator dowhich will turn black if exposed to temperatures above 143 °C. Apply the sticker to the bottom of the data logger itself (not the thermal shield), remove the battery for safety, place the data logger into the thermal shield and run the assembly through the proposed temperature program. The first indicator dot on the sticker will turn black at 143 °C. If that happens, the HiTemp140 with thermal shield is not appropriate for the application and we will work to find a solution that is.

BATTERY WARNING: WARNING: FIRE, EXPLOSION, AND SEVERE BURN HAZARD. DO NOT SHORT CIRCUIT, CHARGE, FORCE OVER DISCHARGE, CRUSH, PENETRATE, OR INCINERATE. BATTERY MAY LEAK OR EXPLODE IF HEATED ABOVE 150 °C (302 °F).

| Maximum Exposure<br>Time Chart           | HiTemp140-TS (Flush) |                         | HiTemp140-TS (Vented) |                         |
|--|----------------------|-------------------------|-----------------------|-------------------------|
| Ambient Temperature                      | Exposure Time in Air | Exposure Time in Liquid | Exposure Time in Air  | Exposure Time in Liquid |
| -200 °C (-328 °F)                        | 12 minutes           | N/A                     | 14 minutes            | N/A                     |
| -180 °C (-292 °F)                        | 13 minutes           | N/A                     | 15 minutes            | N/A                     |
| -160 °C (-256 °F)                        | 15 minutes           | N/A                     | 16 minutes            | N/A                     |
| -140 °C (-220 °F)                        | 17 minutes           | N/A                     | 18 minutes            | N/A                     |
| -120 °C (-184 °F)                        | 19 minutes           | N/A                     | 21 minutes            | N/A                     |
| -100 °C (-148 °F)                        | 22 minutes           | N/A                     | 24 minutes            | N/A                     |
| -80 °C (-112 °F)                         | 27 minutes           | N/A                     | 30 minutes            | N/A                     |
| -60 °C (-76 °F)                          | 37 minutes           | 22 minutes              | 42 minutes            | 25 minutes              |
| -40 °C to +140 °C<br>(-40 °F to +284 °F) | Indefinitely         | Indefinitely            | Indefinitely          | Indefinitely            |
| 150 °C (302 °F)                          | 59 minutes           | 34 minutes              | 66 minutes            | 40 minutes              |
| 160 °C (320 °F)                          | 51 minutes           | 29 minutes              | 57 minutes            | 34 minutes              |
| 170 °C (338 °F)                          | 43 minutes           | 25 minutes              | 48 minutes            | 29 minutes              |
| 180 °C (356 °F)                          | 37 minutes           | 23 minutes              | 42 minutes            | 26 minutes              |
| 190 °C (374 °F)                          | 34 minutes           | 20 minutes              | 38 minutes            | 23 minutes              |
| 200 °C (392 °F)                          | 31 minutes           | 18 minutes              | 34 minutes            | 21 minutes              |
| 210 °C (410 °F)                          | 29 minutes           | 17 minutes              | 32 minutes            | 19 minutes              |
| 220 °C (428 °F)                          | 27 minutes           | 16 minutes              | 30 minutes            | 18 minutes              |
| 230 °C (446 °F)                          | 25 minutes           | 15 minutes              | 27 minutes            | 17 minutes              |
| 240 °C (464 °F)                          | 23 minutes           | 14 minutes              | 26 minutes            | 16 minutes              |
| 250 °C (482 °F)                          | 22 minutes           | 13 minutes              | 24 minutes            | 15 minutes              |

# ORDERING INFORMATION

| MODEL                   | DESCRIPTION   | Enclosure       |
|-------------------------|---|-----------------|
| HITEMP140-5-TSK         | HiTemp140-5.25 inch data logger and thermal shield                      | Flush or Vented |
| HITEMP140-7-TSK         | HiTemp140-7 inch data logger and thermal shield                         | Flush or Vented |
| IFC400                  | Docking station with USB cable, software and manual                     |                 |
| IFC406                  | 6 Port, Multiplexer docking station with USB cable, software and manual |                 |
| ER1425S-HT              | Replacement battery for the HiTemp140                                   |                 |
| Calibration Certificate | Calibration Certificate available for data logger                       |                 |







<sup>\*\*</sup>Other probe lengths up to 7" available.